



PostureFit Lift Recline Chair

Configurable Pressure Care Seating Solution

Prescriber's Name:

Phone Number:

Email:

Funding Body:

Client's Name:

Phone Number:

Address:

Part Number	Size	Seat Depth	Seat Width	Seat-to-floor Height	Backrest Height	SWL
■ CHP198240 ■ CHP198241	Petite	435mm D1 495mm D2	485mm W1 535mm W2	450mm	700mm	160kg
■ CHP198250 ■ CHP198251	Small/Medium (S/M)	475mm D1 525mm D2	485mm W1 535mm W2	490mm	740mm	160kg
■ CHP198260 ■ CHP198261	Medium/Large Tall (M/LT)	475mm D1 525mm D2	535mm W1 585mm W2	490mm	790mm	160kg

Seat Cushion



Pressure Reducing Foam
(Standard) [PR1](#)



After Market Cushion
Petite - 15 x 15in [PR2](#)
S/M - 15 x 15in [PR2](#)
M/LT - 17 x 17in [PR2](#)

Backrest Cushion



Waterfall Backrest
[PO1](#) (Standard)



Kyphosis Backrest (Standard)
[PO2](#) (Standard)



Classic Style Backrest [PO3](#)
Only available in Black while stock lasts
S/M - CHP198290
M/LT - CHP198291



Advance Immersion Backrest [PO4](#)
Petite Black - CHP198293
Petite Stone - CHP198294
S/M Black - CHP198295
S/M Stone - CHP198297
M/LT Black - CHP198296
M/LT Stone - CHP198298

Accessories (Items highlighted in blue)



Channel Legrest [CL](#)
All Sizes Black - CHP198320



Lateral Support Blocks [LS](#)
All Sizes Black - CHP198315



Profiled Headrest [PH](#)
Petite + S/M Black - CHP198310
Petite + S/M Stone - CHP198312
M/LT Black - CHP198311
M/LT Stone - CHP198313

Handset Control Location

Right Hand Side [HC1](#)

Left Hand Side [HC2](#)

Attendant Position - CHP198325 [HC3](#)

CHAIR ASSESSMENT AND SITE INSPECTION CHECK LIST

After assessment please fill out the below form prior to ordering the chair

External Movement

Alternative entry access points:

Are there steps? Yes No If yes, How many?

Is the property accessible easily? Yes No Comments:

Please give extra information if any:

Internal Movement

Are the door and hallway wide enough to move the Chair? Yes No

Does the chair have to go upstairs? Yes No

Is there a lift? Yes No If yes please mention the lift dimensions (L x W x H):

Is the location and space adequate for the chair? Yes No

Is there a power outlet close to the chairs desired location? Yes No

Additional Information: